


VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

AUG 29 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130130
Name		Jeremy Conellings	ID# 6715 Date 08.26.09
A	Agency	VPD	Phone # 835 4560
Instrument Location		VPD Jail	
Alco S/N		x124630	Target Value .078 High Pressure 600
B	Alco Test Value Average	.079	.079
	1 st Alco		2 nd Alco
Signature			271
		(OVER)	08/31/09

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (Do not write in the area below)
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
 - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
 - (3) I am the Scientific Director of the State Breath Alcohol Program.
 - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
 - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification
 - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz

Scientific Director


State Breath Alcohol Program

Subscribed and sworn before me this 17th day of Sept, 2009.

Carolyn M. Noland

Notary Public, State of Alaska

Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130130 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130130

AUGUST 26, 2009 ✓ *CSB 9/1/09*

OPERATOR'S NAME:
CONKLING/JEREMY/V
OPERATOR'S NUMBER: 6715
SUBJECT'S LAST NAME:
UOC
SUBJECT'S FIRST NAME/MI :
UOC

O.L. #: UOC
DEPT/AGENCY: UDZ1
CASE/REPORT: UOC
TEST TYPE: U
ALCO TARGET VALUE: .078 ✓
ALCO S/N: X124630

— BREATH ANALYSIS —

.078 ADJUSTED FOR 30.14 in		
ALCO TARGET	.078	12:24
BLANK TEST	.000	12:25
INTERNAL STANDARD	VERIFIED	12:25
ALCO TO 30.14 in	.079	12:25
BLANK TEST	.000	12:26
SUBJECT SAMPLE	.000	12:26
BLANK TEST	.000	12:27
ALCO TO 30.14 in	.079	12:27
BLANK TEST	.000	12:28

CSB 9/1/09

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130130

AUGUST 26, 2009 ✓
TIME 12:22

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
NUMBERS
SAMPLE CHAMBER: 49c
BURNIN TUBE: 41c
BAROMETER: 30.14 in
FLOW DETECTOR: OKAY

PUMP
PUMP SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST
!"#\$%&'()*+,-./0123456789:;<=?@ABCDEF6
HIJKLMNPOQRSTUVWXYZ[\]^_`abcdefgijklmno
pqrstuvwxyz{|}~